


DECLARATION OF CONFORMITY

| | | |
|--|---|--|
| LEGAL MANUFACTURER | Zest Anchors, LLC 2875 Loker Avenue East Carlsbad, CA 92010 USA | |
| AUTHORIZED REPRESENTATIVE | MDSS GmbH Schiffgraben 41 30175 Hannover, GERMANY SRN#: DE-AR-000005430 | |
| | MDSS CH GmbH Laurenzenvorstadt 61 5000 Aarau Switzerland | |
| RESPONSIBLE PERSON | MDSS-UK RP Limited Parkway House, Palatine Rd, Northenden, Wythenshawe, Manchester M22 4DB United Kingdom | |
| RESPONSIBILITY | This declaration of conformity is issued under the sole responsibility of Zest Anchors, LLC. | |
| PRODUCT(S) | Dust Cabinets (refer to attached product schedule) | |
| CLASSIFICATION | Low Voltage Equipment, Non-Medical Device | |
| INTENDED PURPOSE | As an apparatus with a mains cord intended for countertop use in a dental laboratory for professional use | |
| DECLARATION | We, the manufacturer, Zest Anchors, LLC, declare that the devices covered by this present declaration are in conformity with the applicable provisions for 2014/35/EU (Low Voltage Directive) and 2014/30/EU (EMC Directive). | |
| STANDARDS/COMMON SPECIFICATIONS APPLIED | EN 61010-1 | Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 1: General requirements |
| | IEC/EN 62471 | Photobiological safety of lamps and lamp systems |
| PLACE OF ISSUE | Carlsbad, CA, USA | |
| SIGNED |  <div style="float: right; text-align: right;"> <p>I approve this document Nov 17, 2025 1:11:25 PM PST</p> <p>Nov 17, 2025</p> </div> | |
| | <p style="text-align: center;">0D845160C0E7494592214AB66631083B</p> <hr style="border: 0; border-top: 1px solid black; margin: 5px 0;"/> <div style="display: flex; justify-content: space-between;"> Maleata Hall, Director Regulatory Affairs Date </div> | |

| Revision | Description of Change | Justification | Approval Date | Reference |
|----------|----------------------------|--|---------------|--------------------------|
| Rev A | New document creation | The change is administrative, not affecting design or intended purpose, therefore, is not a significant change under MDCG-2020-3 and is permitted under MDR. | 02/13/25 | Guidance Doc MDCG-2020-3 |
| Rev B | Update logo and formatting | The change is administrative, not affecting design or intended purpose, therefore, is not a significant change under MDCG-2020-3 and is permitted under MDR. | Nov 17, 2025 | Guidance Doc MDCG-2020-3 |

| Part Number | Description |
|-------------|-------------------------------|
| 91010 | MicroCab+ Dust Cabinet (120V) |
| 91354 | MicroCab+ Dust Cabinet (230V) |
| 93817 | MicroCab+ Shield (10 Pack) |
| 49055 | MicroCab+ Filter (CA3915) |
| 91556 | MacroCab+ Dust Cabinet (120V) |
| 91740 | MacroCab+ Dust Cabinet (220V) |
| 93960 | MacroCab+ Shield (10 Pack) |
| 91557 | MacroCab+ Filter (CA7598) |
| 91518 | Handguard Kit (2 Pack) |