

Product description

Alcohol Free Surface
Disinfectant Spray and Wipe




Date of revision: 05/10/20 v2.4

Microorganism	Strain	Contact Time	Condition	Log Reduction	Lab / Report No
EN 14885 : 2015 Chemical disinfectants and antiseptics. Application of European Standards for chemical disinfectants and antiseptics. This European Standard specifies the below European Standards to which products have to conform in order to support the claims for microbicidal activity such as bactericidal etc. It is applicable to products to be used in the area of human medicine.					
BS EN 13727:2012 + A1:2013 Chemical disinfectants and antiseptics – Quantitative suspension test for the evaluation of bactericidal activity in the medical area – Test method and requirements (phase 2, step 1).					
Pseudomonas aeruginosa	ATCC 15442	30 seconds	Dirty	>5.30	MelBec Microbiology Ltd Lab ref: 773.3
Staphylococcus aureus	ATCC 6538	30 seconds	Dirty	>5.34	MelBec Microbiology Ltd Lab ref: 773.3
Enterococcus hirae	ATCC 10541	30 seconds	Dirty	>5.46	MelBec Microbiology Ltd Lab ref: 773.3
VRE	NCTC 2201	1 minute	Dirty	>5.11	MelBec Microbiology Ltd Lab ref: 2281a
MRSA	NCTC 10442	1 minute	Dirty	>5.09	MelBec Microbiology Ltd Lab ref: 2281a
Legionella pneumophila	NCTC 11192	5 minutes	Dirty	>5.52	MelBec Microbiology Ltd Lab ref: 2281a
BS EN 13624:2003 Chemical disinfectants and antiseptics – Quantitative suspension test for the evaluation of fungicidal or yeasticidal activity in the medical area – Test method and requirements (phase 2, step 1).					
Aspergillus niger [®]	NCPF 2275	30 seconds	Dirty	>4.05	Abbott Analytical Certificate No 11B 061MF TOP 22/02/11
Candida albicans	NCPF 3179	30 seconds	Dirty	>4.19	Abbott Analytical Certificate No 11B 061MF TOP 22/02/11

BS EN 14348:2005

Chemical disinfectants and antiseptics. Quantitative suspension test for the evaluation of mycobactericidal activity of chemical disinfectants in the medical area including instrument disinfectants. Test methods and requirements (phase 2, step 1)

Mycobacterium avium	ATCC 15769	2 min	Clean	++	BluScientific Glasgow Certificate dated 14-11-2008 BS-TDL-001

BS EN 13704:2002

Chemical disinfectants. Quantitative suspension test for the evaluation of sporicidal activity of chemical disinfectants used in food, industrial, domestic and institutional areas. Test method and requirements (phase 2, step 1)

Clostridium difficile (C-diff)	NCTC 11209	5 mins	Clean	4-86	Abbott Analytical Certificate No 08L022cd TOP

BS EN 14476:2013 + A2:2019

Chemical disinfectants and antiseptics – Quantitative suspension test for the evaluation of virucidal activity in the medical area - Test method and requirements (Phase 2/Step 1)

Vaccinia virus	VR-1549 Elstree strain (P7)	2 mins	Dirty	>4	BluTest Laboratories Ltd. Lab code BT-TOP-23 A1

Efficacy Test Report Certificate



Report for BS EN 13727:2012 + A1:2013

Chemical disinfectants and antiseptics – Quantitative suspension test for the evaluation of bactericidal activity in the medical area – Test method and requirements (phase 2, step 1).

Topdental
12 Ryefield Way,
Silsden,
West Yorkshire.
BD20 0EF

Report Date: 24.04.16
Prepared By: Dawn Mellors
Date Tested: 19.04.16 & 20.04.16

01535 652750

science@topdental.co.uk

Contact Name: Will Howell
Purchase Order Number: NA

Test Details:

	Product and Test Criteria	Product and Test Details
Product Details	Lab reference	773.3
	Name of Product(s)	Ready to Use Alcohol Free Surface Cleaner
	Manufacturer	Topdental
	Batch Code	07/04/16
	Storage Conditions	Ambient
	Active Substance	-
	Product Diluent	Not required. Product supplied as a Ready to Use product with concentration adjusted to compensate for dilution during testing.
Experimental Conditions	Method	BS EN 13727:2012 + A1:2013
	Neutraliser	Membrane Filtration for <i>Pseudomonas aeruginosa</i> . BU Broth for <i>Staphylococcus aureus</i> and <i>Enterococcus hirae</i> .
	Plating Method	Pour Plate using Tryptone Soy Agar
	Test Product Dilutions	Ready to Use
	Identification of the bacterial strains	<i>Pseudomonas aeruginosa</i> ATCC 15442 <i>Staphylococcus aureus</i> ATCC 6538 <i>Enterococcus hirae</i> ATCC 10541
	Test Temperature	20°C ± 1°C
	Contact Times	30 second
	Interfering Substance	Dirty Conditions: 3.0g/l Bovine Albumin and 3ml/l sheep erythrocytes
	Incubation Temperature	36°C
	Product Dilutions	Client only requested testing on the product at one dilution.

Unit 1, Envair House, York Avenue, Haslingden, Rossendale, BB4 4HX
info@melbecmicrobiology.co.uk Tel – 01706 214 492 melbecmicrobiology.co.uk

Registered in England. Company Number – 8604622 VAT Reg – 167 8579 45

Efficacy Test Report Certificate



Requirements of the Standard:

The test product shall demonstrate at least a 5 decimal log (lg) reduction when tested in accordance with this standard.

Results:

Test Suspension: (N & N₀)

<i>Pseudomonas aeruginosa</i> ATCC 15442	N	V _c 1	V _c 2	xwm = 2.8 x 10 ⁸ lgN = 8.45 N ₀ = 2.8 x 10 ⁷ lg N ₀ = 7.45 7.18 ≤ 7.45 ≤ 7.70
	10 ⁻⁷	27	29	
<i>Staphylococcus aureus</i> ATCC 6538	N	V _c 1	V _c 2	xwm = 3.1 x 10 ⁸ lgN = 8.49 N ₀ = 3.1 x 10 ⁷ lg N ₀ = 7.49 7.18 ≤ 7.49 ≤ 7.70
	10 ⁻⁷	30	32	
<i>Enterococcus hirae</i> ATCC 10541	N	V _c 1	V _c 2	xwm = 4.1 x 10 ⁸ lgN = 8.61 N ₀ = 4.1 x 10 ⁷ lg N ₀ = 7.61 7.18 ≤ 7.61 ≤ 7.70
	10 ⁻⁷	43	38	

Test (N_a and lgR)

N_a is the number of survivors per ml in the test mixture at the end of the contact time and before neutralisation.

lgR is the log reduction.

Organism	Conc. of product %	V _c Test 1	V _c Test 2	N _a x 10	lg N _a	lgR
<i>Pseudomonas aeruginosa</i>	100% (80%)	0-1	0-1	<140	<2.15	>5.30
	-	-	-	-	-	-
<i>Staphylococcus aureus</i>	100% (80%)	0-1	0-1	<140	<2.15	>5.34
	-	-	-	-	-	-
<i>Ent. hirae</i>	100% (80%)	0-1	0-1	<140	<2.15	>5.46
	-	-	-	-	-	-

Unit 1, Envair House, York Avenue, Haslingden, Rossendale, BB4 4HX
info@melbecmicrobiology.co.uk Tel – 01706 214 492 melbecmicrobiology.co.uk

Registered in England. Company Number – 8604622 VAT Reg – 167 8579 45

Efficacy Test Report Certificate



Validation and Controls:

Pseudomonas aeruginosa

Validation Suspension (N _{v0})			Experimental Conditions Control (A)		
V _{c1}	50	$\bar{x} = 4.35$ $\times 10^1$	V _{c1}	39	$\bar{x} = 3.6 \times$ 10^1
V _{c2}	37		V _{c2}	33	
30 ≥ N _{v0} ≤ 160			$\bar{x}$ of A is ≥ 0.5x $\bar{x}$ of N _{v0} Yes		

Validation Suspension (N _v)			Neutraliser Control (B)		
V _c 1	50	x̄ = 4.35 x 10 ⁴	V _c 1	42	x̄ = 3.8 x 10 ³
V _c 2	37		V _c 2	34	
3.0 x 10 ⁴ ≥ N _v ≥ 1.6 x 10 ⁵			x̄ of B is ≥ 0.5x x̄ of N _v B Yes		

Staphylococcus aureus

Validation Suspension (N _{v0})			Experimental Conditions Control (A)		
V _{c1}	69	$\bar{x} = 7.45 \times 10^1$	V _{c1}	72	$\bar{x} = 7.75 \times 10^1$
V _{c2}	80		V _{c2}	83	
30 ≥ N _{v0} ≤ 160			$\bar{x}$ of A is ≥ 0.5x $\bar{x}$ of N _{v0} Yes		

Validation Suspension (N _{vB})			Neutraliser Control (B)		
V _{c1}	92	$\bar{x} = 8.5 \times 10^4$	V _{c1}	78	$\bar{x} = 8.7 \times 10^3$
V _{c2}	78		V _{c2}	96	
$3.0 \times 10^4 \geq N_{vB} \leq 1.6 \times 10^5$			$\bar{x}$ of B is $\geq 0.5 \times \bar{x}$ of N _{vB} Yes		

Enterococcus hirae

Validation Suspension (N _{v0})			Experimental Conditions Control (A)		
V _{c1}	123	$\bar{x} = 1.16 \times 10^2$	V _{c1}	95	$\bar{x} = 9.6 \times 10^1$
V _{c2}	109		V _{c2}	77	
30 ≥ N _{v0} ≤ 160			$\bar{x}$ of A is ≥ 0.5x $\bar{x}$ of N _{v0} Yes		

Validation Suspension (N _{VB})			Neutraliser Control (B)		
V _{C1}	119	$\bar{x} = 1.1 \times 10^5$	V _{C1}	90	$\bar{x} = 8.4 \times 10^3$
V _{C2}	101		V _{C2}	78	
$3.0 \times 10^4 \geq N_{VB} \leq 1.6 \times 10^5$			$\bar{x}$ of B is $\geq 0.5 \times \bar{x}$ of N _{VB} Yes		

Unit 1, Envair House, York Avenue, Haslingden, Rossendale, BB4 4HX
info@melbecmicrobiology.co.uk Tel - 01706 214 492 melbecmicrobiology.co.uk

Registered in England. Company Number - 8604622 VAT Reg - 167 8579 45

Efficacy Test Report Certificate



Method Validation (C):

<i>Pseudomonas aeruginosa</i>			<i>Staphylococcus aureus</i>			<i>Ent hirae</i>		
V _c 1	29	$\bar{x}$ = 3.2 x 10 ¹	V _c 1	0	$\bar{x}$ = <10	V _c 1	111	$\bar{x}$ = 5.9 x 10 ²
V _c 2	35		V _c 2	0		V _c 2	105	
$\bar{x}$ of C is $\geq 0.5 \times \bar{x}$ of N _v 0 Yes			$\bar{x}$ of C is $\geq 0.5 \times \bar{x}$ of N _v 0 NO			$\bar{x}$ of C is $\geq 0.5 \times \bar{x}$ of N _v 0 Yes		

Conclusion:

The test product has met the requirements (with a 30 second contact time) as specified in EN 13727 (5 log reduction) in dirty conditions as a Ready to Use product.

Please note that the neutralisation validation for the *Staphylococcus aureus* has not been successful using either neutralising diluent or membrane filtration. A full dilution of the neutraliser / product mix was carried out and no growth was observed at any dilution. When neutralisation is a problem and an organism has not been killed it will usually be observed on the 10^{-2} dilution as the active ingredient has been diluted enough to allow growth.

Report Authorised By:

Dawn Mellors

Technical Director

This is the end of the test report

Unit 1, Envair House, York Avenue, Haslingden, Rossendale, BB4 4HX
info@melbecmicrobiology.co.uk Tel – 01706 214 492 melbecmicrobiology.co.uk

Registered in England. Company Number – 8604622 VAT Reg – 167 8579 45

Efficacy Test Report Certificate



Report for BS EN 13727:2012 + A1:2013

Chemical disinfectants and antiseptics – Quantitative suspension test for the evaluation of bactericidal activity in the medical area – Test method and requirements (phase 2, step 1).

Topdental
12 Ryefield Way,
Silsden,
West Yorkshire.
BD20 0EF
01535 652750

Report Date: 12.03.17
Prepared By: Dawn Mellors
Date Tested: 23/02/17 & 24/02/17

science@topdental.co.uk

Contact Name: Will Howell
Purchase Order Number: NA

Test Details:

	Product and Test Criteria	Product and Test Details
Product Details	Lab reference	2281a
	Name of Product(s)	WH-10 Alcohol Free
	Manufacturer	Topdental
	Batch Code	20/02/17
	Storage Conditions	Ambient
	Active Substance	-
	Product Diluent	Not required. Product supplied as a Ready to Use product with concentration adjusted to compensate for dilution during testing.
Experimental Conditions	Method	BS EN 13727:2012 + A1:2013
	Neutraliser	Membrane Filtration
	Plating Method	Pour Plate using Tryptone Soy Agar
	Test Product Dilutions	Ready to Use
	Identification of the bacterial strains	<i>E. coli</i> ATCC 10536 <i>Salmonella</i> Typhimurium NCTC 12023 VRE NCTC 2201 MRSA NCTC 10442 <i>Serratia marcescens</i> <i>Listeria monocytogenes</i> NCTC 7973 <i>Legionella pneumophila</i> NCTC 11192
	Test Temperature	20°C ± 1°C
	Contact Times	1 min & 5 min
	Interfering Substance	Dirty Conditions: 3.0g/l Bovine Albumin & Sheep erythrocytes

Unit 1, Envair House, York Avenue, Haslingden, Rossendale, BB4 4HX
info@melbecmicrobiology.co.uk Tel – 01706 214 492 melbecmicrobiology.co.uk

Registered in England. Company Number – 8604622 VAT Reg – 167 8579 45

Efficacy Test Report Certificate



	Incubation Temperature	36°C
Deviations	Product Dilutions	Client only requested testing on the product at one dilution. Timepoints chosen by client. Test organisms chosen by the client.

Requirements of the Standard:

The test product shall demonstrate at least a 5 decimal log (lg) reduction when tested in accordance with the standard.

Results:

Test Suspension: (N & N₀)

<i>E. coli</i>	N	V _c 1	V _c 2	$\dot{x}_{wm} = 1.87 \times 10^8$ lgN = 8.27 $N_0 = 1.87 \times 10^7$ lg N ₀ = 7.27 $7.18 \leq 7.27 \leq 7.70$
	10 ⁻⁷	20	23	
	10 ⁻⁶	170	199	
<i>Salmonella</i> <i>Typhimurium</i>	N	V _c 1	V _c 2	$\dot{x} = 2.28 \times 10^8$ lgN = 8.36 $N_0 = 2.28 \times 10^7$ lg N ₀ = 7.36 $7.18 \leq 7.36 \leq 7.70$
	10 ⁻⁷	19	18	
	10 ⁻⁶	227	238	
VRE	N	V _c 1	V _c 2	$\dot{x} = 1.84 \times 10^8$ lgN = 8.26 $N_0 = 1.84 \times 10^7$ lg N ₀ = 7.26 $7.18 \leq 7.26 \leq 7.70$
	10 ⁻⁷	16	15	
	10 ⁻⁶	179	194	
MRSA	N	V _c 1	V _c 2	$\dot{x} = 1.53 \times 10^8$ lgN = 8.18 $N_0 = 1.53 \times 10^7$ lg N ₀ = 7.18 $7.18 \leq 7.24 \leq 7.70$
	10 ⁻⁷	<14	16	
	10 ⁻⁶	150	155	
<i>Legionella</i> <i>pneumophila</i>	N	V _c 1	V _c 2	$\dot{x} = 4.65 \times 10^8$ lgN = 8.67 $N_0 = 4.65 \times 10^7$ lg N ₀ = 7.67 $7.18 \leq 7.67 \leq 7.70$
	10 ⁻⁷	49	44	
	10 ⁻⁶	>300	>300	
<i>Listeria</i> <i>monocytogenes</i>	N	V _c 1	V _c 2	$\dot{x} = 4.65 \times 10^8$ lgN = 8.67 $N_0 = 4.65 \times 10^7$ lg N ₀ = 7.67 $7.18 \leq 7.67 \leq 7.70$
	10 ⁻⁷	47	46	
	10 ⁻⁶	>300	>300	
<i>Serratia marcescens</i>	N	V _c 1	V _c 2	$\dot{x} = 2.64 \times 10^8$ lgN = 8.42 $N_0 = 2.64 \times 10^7$ lg N ₀ = 7.42 $7.18 \leq 7.42 \leq 7.70$
	10 ⁻⁷	24	34	
	10 ⁻⁶	254	268	
<i>Klebsiella</i> <i>pneumoniae</i>	N	V _c 1	V _c 2	$\dot{x} = 3.8 \times 10^8$ lgN = 8.58 $N_0 = 3.8 \times 10^7$ lg N ₀ = 7.58 $7.18 \leq 7.58 \leq 7.70$
	10 ⁻⁷	33	43	
	10 ⁻⁶	>300	>300	

Unit 1, Envair House, York Avenue, Haslingden, Rossendale, BB4 4HX
 info@melbecmicrobiology.co.uk Tel – 01706 214 492 melbecmicrobiology.co.uk

Registered in England. Company Number – 8604622 VAT Reg – 167 8579 45

Efficacy Test Report Certificate



Test (N_a and IgR)

N_a is the number of survivors per ml in the test mixture at the end of the contact time and before neutralisation.

IgR is the log reduction.

1 min

Organism	Conc. of product %	V_c Test 1	V_c Test 2	N_a $\times 10$	Ig N_a	IgR
<i>E. coli</i>	100% (80%)	0-1	0-1	<140	<2.15	>5.12
	-	-	-	-	-	-
<i>Salmonella</i> Typhimurium	100% (80%)	0-1	0-1	<140	<2.15	>5.21
	-	-	-	-	-	-
VRE	100% (80%)	0-1	0-1	<140	<2.15	>5.11
	-	-	-	-	-	-
MRSA	100% (80%)	0-1	0-1	<140	<2.15	>5.09
	-	-	-	-	-	-
<i>Leg pneumophila</i>	100% (80%)	60-4	74-4	6.7×10^5	5.83	1.84
	-	-	-	-	-	-
<i>Klebsiella pneumoniae</i>	100% (80%)	0-1	0-1	<140	<2.15	>5.43
	-	-	-	-	-	-
<i>Listeria monocytogenes</i>	100% (80%)	0-1	0-1	<140	<2.15	>5.52
	-	-	-	-	-	-
<i>Serratia marcescens</i>	100% (80%)	0-1	0-1	<140	<2.15	>5.26
	-	-	-	-	-	-

5 min

Organism	Conc. of product %	V_c Test 1	V_c Test 2	N_a $\times 10$	Ig N_a	IgR
<i>E. coli</i>	100% (80%)	0-1	0-1	<140	<2.15	>5.12
	-	-	-	-	-	-
<i>Salmonella</i> Typhimurium	100% (80%)	0-1	0-1	<140	<2.15	>5.21
	-	-	-	-	-	-
VRE	100% (80%)	0-1	0-1	<140	<2.15	>5.11
	-	-	-	-	-	-
MRSA	100% (80%)	0-1	0-1	<140	<2.15	>5.09
	-	-	-	-	-	-
<i>Leg pneumophila</i>	100% (80%)	5-1	6-1	<140	<2.15	>5.52
	-	-	-	-	-	-
<i>Klebsiella pneumoniae</i>	100% (80%)	0-1	0-1	<140	<2.15	>5.43
	-	-	-	-	-	-
<i>Listeria monocytogenes</i>	100% (80%)	0-1	0-1	<140	<2.15	>5.52
	-	-	-	-	-	-
<i>Serratia marcescens</i>	100% (80%)	0-1	0-1	<140	<2.15	>5.26
	-	-	-	-	-	-

Unit 1, Envair House, York Avenue, Haslingden, Rossendale, BB4 4HX
 info@melbecmicrobiology.co.uk Tel – 01706 214 492 melbecmicrobiology.co.uk

Registered in England. Company Number – 8604622 VAT Reg – 167 8579 45

Efficacy Test Report Certificate



Validation and Controls:

E. coli

Validation Suspension (N _{vo})			Experimental Conditions Control (A)			Neutraliser Control (B)			Method Validation (C):		
V _{c1}	47	$\bar{x} = 4.15 \times 10^1$	V _{c1}	59	$\bar{x} = 6.1 \times 10^1$	V _{c1}	59	$\bar{x} = 5.2 \times 10^1$	V _{c1}	41	$\bar{x} = 4.5 \times 10^1$
V _{c2}	36		V _{c2}	63		V _{c2}	45		V _{c2}	49	
30 ≥ N _{vo} ≤ 160			$\bar{x}$ of A is ≥ 0.5x $\bar{x}$ of N _{vo} Yes			$\bar{x}$ of B is ≥ 0.5x $\bar{x}$ of N _{vo} Yes			$\bar{x}$ of C is ≥ 0.5x $\bar{x}$ of N _{vo} Yes		

Salmonella Typhimurium

Validation Suspension (N _{vo})			Experimental Conditions Control (A)			Neutraliser Control (B)			Method Validation (C):		
V _{c1}	50	$\bar{x} = 5.85 \times 10^1$	V _{c1}	58	$\bar{x} = 6.9 \times 10^1$	V _{c1}	57	$\bar{x} = 5.3 \times 10^1$	V _{c1}	39	$\bar{x} = 4.2 \times 10^1$
V _{c2}	67		V _{c2}	80		V _{c2}	49		V _{c2}	45	
30 ≥ N _{vo} ≤ 160			$\bar{x}$ of A is ≥ 0.5x $\bar{x}$ of N _{vo} Yes			$\bar{x}$ of B is ≥ 0.5x $\bar{x}$ of N _{vo} Yes			$\bar{x}$ of C is ≥ 0.5x $\bar{x}$ of N _{vo} Yes		

Klebsiella pneumoniae

Validation Suspension (N _{vo})			Experimental Conditions Control (A)			Neutraliser Control (B)			Method Validation (C):		
V _{c1}	82	$\bar{x} = 9.0 \times 10^1$	V _{c1}	63	$\bar{x} = 6.7 \times 10^1$	V _{c1}	77	$\bar{x} = 7.45 \times 10^1$	V _{c1}	64	$\bar{x} = 7.8 \times 10^1$
V _{c2}	98		V _{c2}	72		V _{c2}	71		V _{c2}	92	
30 ≥ N _{vo} ≤ 160			$\bar{x}$ of A is ≥ 0.5x $\bar{x}$ of N _{vo} Yes			$\bar{x}$ of B is ≥ 0.5x $\bar{x}$ of N _{vo} Yes			$\bar{x}$ of C is ≥ 0.5x $\bar{x}$ of N _{vo} Yes		

Listeria monocytogenes

Validation Suspension (N _{vo})			Experimental Conditions Control (A)			Neutraliser Control (B)			Method Validation (C):		
V _{c1}	127	$\bar{x} = 1.09 \times 10^2$	V _{c1}	126	$\bar{x} = 1.39 \times 10^2$	V _{c1}	107	$\bar{x} = 1.01 \times 10^2$	V _{c1}	129	$\bar{x} = 1.37 \times 10^2$
V _{c2}	91		V _{c2}	151		V _{c2}	95		V _{c2}	145	
30 ≥ N _{vo} ≤ 160			$\bar{x}$ of A is ≥ 0.5x $\bar{x}$ of N _{vo} Yes			$\bar{x}$ of B is ≥ 0.5x $\bar{x}$ of N _{vo} Yes			$\bar{x}$ of C is ≥ 0.5x $\bar{x}$ of N _{vo} Yes		

VRE

Validation Suspension (N _{vo})			Experimental Conditions Control (A)			Neutraliser Control (B)			Method Validation (C):		
V _{c1}	39	$\bar{x} = 3.7 \times 10^1$	V _{c1}	54	$\bar{x} = 4.8 \times 10^1$	V _{c1}	45	$\bar{x} = 4.3 \times 10^1$	V _{c1}	39	$\bar{x} = 3.3 \times 10^1$
V _{c2}	35		V _{c2}	42		V _{c2}	48		V _{c2}	27	
30 ≥ N _{vo} ≤ 160			$\bar{x}$ of A is ≥ 0.5x $\bar{x}$ of N _{vo} Yes			$\bar{x}$ of B is ≥ 0.5x $\bar{x}$ of N _{vo} Yes			$\bar{x}$ of C is ≥ 0.5x $\bar{x}$ of N _{vo} Yes		

MRSA

Validation Suspension (N _{vo})			Experimental Conditions Control (A)			Neutraliser Control (B)			Method Validation (C):		
V _{c1}	38	$\bar{x} = 3.65 \times 10^1$	V _{c1}	27	$\bar{x} = 2.5 \times 10^1$	V _{c1}	28	$\bar{x} = 2.75 \times 10^1$	V _{c1}	20	$\bar{x} = 1.85 \times 10^1$
V _{c2}	35		V _{c2}	23		V _{c2}	27		V _{c2}	17	
30 ≥ N _{vo} ≤ 160			$\bar{x}$ of A is ≥ 0.5x $\bar{x}$ of N _{vo} Yes			$\bar{x}$ of B is ≥ 0.5x $\bar{x}$ of N _{vo} Yes			$\bar{x}$ of C is ≥ 0.5x $\bar{x}$ of N _{vo} Yes		

Unit 1, Envair House, York Avenue, Haslingden, Rossendale, BB4 4HX

info@melbecmicrobiology.co.uk Tel – 01706 214 492 melbecmicrobiology.co.uk

Registered in England. Company Number – 8604622 VAT Reg – 167 8579 45

Validation Suspension (N_{v0})			Experimental Conditions Control (A)			Neutraliser Control (B)			Method Validation (C):		
Vc1	38	$\dot{x} = 4.15 \times 10^1$	Vc1	27	$\dot{x} = 2.5 \times 10^1$	Vc1	32	$\dot{x} = 3.45 \times 10^1$	Vc1	41	$\dot{x} = 3.7 \times 10^1$
Vc2	45		Vc2	23		Vc2	37		Vc2	33	
$30 \geq N_{v0} \leq 160$			$\dot{x}$ of A is $\geq 0.5x$ $\dot{x}$ of N_{v0} Yes			$\dot{x}$ of B is $\geq 0.5x$ $\dot{x}$ of N_{v0} Yes			$\dot{x}$ of C is $\geq 0.5x$ $\dot{x}$ of N_{v0} Yes		

Validation Suspension (N _{v0})			Experimental Conditions Control (A)			Neutraliser Control (B)			Method Validation (C):		
V _{c1}	71	$\dot{x} = 8.6 \times 10^1$	V _{c1}	60	$\dot{x} = 5.65 \times 10^1$	V _{c1}	51	$\dot{x} = 5.45 \times 10^1$	V _{c1}	89	$\dot{x} = 9.5 \times 10^1$
V _{c2}	81		V _{c2}	53		V _{c2}	58		V _{c2}	101	
30 ≥ N _{v0} ≤ 160			$\dot{x}$ of A is ≥ 0.5x $\dot{x}$ of N _{v0} Yes			$\dot{x}$ of B is ≥ 0.5x $\dot{x}$ of N _{v0} Yes			$\dot{x}$ of C is ≥ 0.5x $\dot{x}$ of N _{v0} Yes		

With a 5 minute contact time the test product met the requirements of the standard (5 log reduction) in dirty conditions as a Ready to Use product for the test organisms.

10/2/20

Technical Director

Unit 1, Envair House, York Avenue, Haslingden, Rossendale, BB4 4HX
info@melbecmicrobiology.co.uk Tel – 01706 214 492 melbecmicrobiology.co.uk

Topdental (Products) Ltd | 12 Ryefield Way Silsden W. Yorks BD20 0EF England UK | Tel +44 (0)1535 652 750 | www.topdental.co.uk

Efficacy Test Report Certificate



Abbott Analytical

Consulting Scientists to the Disinfectant Industry



Certificate of Analysis

Sample(s): One sample of Mix Control
Lot 102777

Received from: Topdental (Products) Ltd, Dental House, Steeton Grove
Industrial Park, Steeton, West Yorkshire, BD20 6TT

Date received: 14 February 2011 **Date tested:** 18 February 2011

Certificate no: 11B.061MF.TOP **Certificate date:** 22 February 2011

Sample ref: 11B/061 **Page:** 1 of 2

Analysis required: EN 13624, Chemical disinfectants and antiseptics -
Quantitative suspension test for the evaluation of
fungicidal and yeasticidal activity of chemical
disinfectants for instruments used in the medical area -
Test method and requirements (phase 2, step 1)

Product stored at: Room temperature

Active substance: Not declared

Test conditions: Dirty

Interfering substance: 3.0g/l bovine albumin +
3.0ml/l sheep erythrocytes

Product test concentration: 8ml in 500ml of water

Product diluent used during test: Sterile hard water 300mg/l CaCO₃

Contact time: 30 seconds

Test temperature: 25°C ± 0.5°C

Neutralising solution: 30g/l polysorbate 80, 3g/l lecithin,
1g/l histidine, 1g/l cysteine;
in phosphate buffer

Incubation temperature: 30°C ± 1°C

Identification of fungal/yeast strain(s) used: *Aspergillus niger* NCPF 2275
Candida albicans NCPF 3179

D C Watson

D C Watson BSc, CBiol, MIBiol, MIFST, ACIEHO
PO Box 95, New Ferry, Wirral, CH62 6HA
Tel: 0151 637 3331 Mob: 07767 871275
email: abbottanalytical@hotmail.co.uk

Efficacy Test Report Certificate

Abbott Analytical

Consulting Scientists to the Disinfectant Industry

22 February 2011

Certificate No: 112.061MF.T02

Page: 2 of 2

Test results:

Test Organism	<i>Aspergillus niger</i>		<i>Candida albicans</i>	
Validation Suspension (Nv.)	Vol 238	Vol 240	Vol 238	Vol 242
	$\bar{x} = 249$		$\bar{x} = 238$	
Experimental Control (A)	Vol 224	Vol 248	Vol 216	Vol 196
	$\bar{x} = 236 \pm 0.5Nv$		$\bar{x} = 206 \pm 0.5Nv$	
Neutraliser Control (B)	Vol 250	Vol 216	Vol 204	Vol 220
	$\bar{x} = 233 \pm 0.5Nv$		$\bar{x} = 212 \pm 0.5Nv$	
Method Validation (C)	Vol 246	Vol 222	Vol 230	Vol 196
	$\bar{x} = 234 \pm 0.5Nv$		$\bar{x} = 213 \pm 0.5Nv$	
Test Suspension	Vol 192	Vol 224	Vol 234	Vol 216
	Vol 31	Vol 24	Vol 16	Vol 23
(N)	$\bar{x} = 2.14 \times 10^6$		$\bar{x} = 2.19 \times 10^6$	
	$\lg N = 7.33$		$\lg N = 7.34$	
(N ₀ = 0.1N)	$\lg N_0 = 6.33$		$\lg N_0 = 6.34$	
Results	Vol 17	Vol 21	Vol 114	Vol 114
(Na)	$10\bar{x} = 190$		$10\bar{x} < 140$	
	$\lg Na = 2.28$		$\lg Na < 2.15$	
(R)	$\lg R = 4.05$		$\lg R > 4.13$	
Pass: $\lg R \geq 4$	PASS		PASS	

Nv = plate count per ml
x = average of Vol 1 and Vol 2

$\bar{x}$ = weighted mean of 2
P = reduction $\lg R = \lg N_0 - \lg Na$

Requirements & Conclusion:

This batch of Mix Control, when diluted 2ml in 500ml of water, passes the requirements of BS 1865 for fungicidal and yeasticidal activity in 30 seconds at 20°C under dirty conditions against the reference organisms detailed.

D. G. Watson

Dr. G. Watson BSc. CIVIL ENGINEER, MSc. MATHS
PO Box 95, New Ferry, Cheshire CH63 4NS
Tel: 0151 617 2221, Mob: 07167 91115
email: g.watson@topdental.co.uk

BluScientific Test Data

Test Report BS EN 14348: 2005. Chemical disinfectants and antiseptics - Quantitative suspension test for the evaluation of mycobactericidal activity of chemical disinfectants in the medical area including instrument disinfectants— Test method and requirements (phase 2, step 1)

Test Laboratory	BluScientific Test Data School of Life Sciences Glasgow Caledonian University GLASGOW G4 0BA
Identification of sample	
Name of the product	MIX0007 Mix Control TM Disinfectant Spray Cartridge
Batch Number	LOT 1024456
Manufacturer	TOPDENTAL (PRODUCTS) LTD, DENTAL HOUSE, STEETON GROVE INDUSTRIAL PARK, STEETON, WEST YORKSHIRE, BD20 6TT
Date of Delivery	14 July 2008
Project code	BS-TDL-001
Storage conditions	4°C and darkness
Active substances	Not known
Test Method and its validation	
Method	Filtration-neutralization
Neutralizer	Lecithin 3g/l, Polysorbate 80 30g/l, sodium thiosulphate 5g/l, L-histidine 1g/l, phosphate buffer 0.0025mol/l, sterilized by autoclave
Experimental Conditions	
Period of analysis	22 October to 11 November 2008
Product diluent used	Sterile hard water
Product test concentrations	7.9 ml to 492 ml distilled water
Appearance product dilutions	Colourless, clear product solution
Contact time	$t = 2 \text{ min} \pm 10 \text{ s}$
Test temperature	$20^{\circ}\text{C} \pm 1^{\circ}\text{C}$
Interfering substance	0.3 g/l bovine albumin
Stability of mixture	Precipitate absent throughout the test
Temperature of incubation	$37^{\circ}\text{C} \pm 1^{\circ}\text{C}$
Identification of strain	<i>Mycobacterium Avium</i> ATCC 15769

Conclusion

According to EN 14348 (2005) the MIX0007 Mix Control TM Disinfectant Spray Cartridge, batch number LOT 1024456, used as received from Topdental *possesses mycobactericidal activity* in 2 min at 20°C under CLEAN conditions (0.3 g/l bovine albumin) for reference strain *Mycobacterium Avium* ATCC 15769

Signed



Dr Chris Woodall, Director
BluScientific Test Data
Glasgow, UK, 14 November 2008

School of Life Sciences, Glasgow Caledonian University, Glasgow G4 0BA, Scotland, UK
T: +44 (0)141 3311285, F: +44 (0)141 3311286, E: life@glasgow.ac.uk, life@glasgow.ac.uk
A: 02045622228, W: www.bluscientific.com


— BluScientific is based in the School of Life Sciences at Glasgow Caledonian University.
Glasgow Caledonian University is a registered charity (company number 0171243)



Efficacy Test Report Certificate



Abbott Analytical



Consulting Scientists to the Disinfectant Industry

24th November 2008

CERTIFICATE OF ANALYSIS.

Samples: One sample of Mix control Hard Surface Spray received from Topdental Products Ltd, Dental House, Steeton Grove Industrial Park, Steeton, West Yorkshire. BD20 6TT 13th November 2008

Certificate No: 08L.022cd.TOP

Page: 1 of 1

Sample Ref: 81 / 022

Analysis Required: BS EN 13704 Quantitative Suspension Test for the Evaluation of Sporicidal activity of chemical disinfectants.

Samples Tested: 19th November 2007

Product stored at room temperature.

Product number - MIX0007

Experimental conditions:

Product test concentrations	- Neat as received.
Interfering substance	- 0.3g/l Bovine serum
Contact time	- 5 mins
Test Temperature	- 20°C $\pm$ 0.5°C
Neutralising solution	- 3% Tween 80, 3% Saponin, 0.1% Histidine, 0.1% Cysteine
Temperature of incubation	- 30°C $\pm$ 1°C

Identification of bacterial strains used - *Clostridium difficile* NCTC 11209



D C Watson

Efficacy Test Report Certificate



Abbott Analytical

Consulting Scientists to the Disinfectant Industry



19th November 2008

Certificate No 8L.022cd.TOP

Page 2 of 2

Test Results

Validation test	Clostridium difficile		
Bacterial suspension	Vc	310,	262
	Nv	2.86×10^3	
Experimental conditions	Vc	224,	270
	A	2.47×10^2	
Neutraliser control	Vc	200,	198
	B	1.99×10^2	
Dilution-neutralisation control	Vc	242,	270
	C	2.56×10^2	
Bacterial Test Suspension	10^{-6}	232	210
	10^{-7}	29	33
	N	2.65×10^6	
Test results			
Neat	Vc	37	
	Na	3700	
	R	7.16×10^4	
Log reduction	4.86		

Vc = Viable Count.

N = Number of cfu/ml of the bacterial test suspension.

Nv = Number of cfu in bacterial suspension.

R = Reduction in viability.

Na = Number of cfu/ml in the test mixture

Conclusion: According to EN13704 this batch of Mix Control Hard Surface Spray when used neat as received possesses satisfactory sporicidal activity in 5 minutes at 20°C under clean conditions (0.3g/l bovine albumin) for the reference organism detailed.

D C Watson

Efficacy Test Report Certificate



Test Report: BS EN 14476:2013 + A2:2019 Chemical disinfectants and antiseptics – Quantitative suspension test for the evaluation of virucidal activity in the medical area- Test method and requirements (Phase 2/Step 1)

Test Laboratory

BluTest Laboratories Ltd

5 Robroyston Oval, Nova Business Park, Glasgow, G33 1AP

Identification of sample

Name of the product	Alcohol Free BT-57
Batch number	2006081
Client	Top Dental Products Limited
Client Address	12 Ryefield Way, Silsden, West Yorkshire, BD20 0EF, UK
Project Code	BT-TOP-23 A1
Date of Delivery	24 June 2020
Storage conditions	Ambient
Active substances	DDAC/ BAC/ PHMB
Appearance	Liquid
Condition upon receipt	Undamaged

Test Method and its validation

Method

1 part interfering substance + 1 part virus suspension + 8 parts biocide were mixed and incubated at the indicated contact temperature for the indicated contact times. Assays were validated by a cytotoxicity control, interference control, neutralisation control and a formaldehyde internal standard.

Neutralisation

Dilution-neutralisation/gel filtration
Eagles Minimum Essential Medium + 5.0% v/v foetal bovine serum at 4°C

Experimental Conditions

Period of analysis	11 August 2020 to 16 August 2020
Product diluents used	Sterile distilled water
Product test concentrations	10.0% v/v; 50.0% v/v; 80.0% v/v
Appearance product dilutions	No changes noted- stable
Appearance in test mixture	Sedimentation observed at 50.0% v/v and 10.0% v/v Turbidity observed at 10.0% v/v
Contact times (minutes)	2 ± 10s
Test temperature	20°C ± 1°C
Interfering substances	3.0 g/l bovine albumin + 3.0 ml/l erythrocytes
Temperature of incubation	37°C ± 1°C + 5% CO ₂
Identification and passage (P) of virus	Vaccinia virus VR-1549 Elstree strain (P7)
Identification and passage (P) of cells	Vero Cells (P 42)

Page 1 of 7

SOP 11000
SOP 8003 EN14476 Vaccinia REPORT TEMPLATE V01
Effective Date: 23 March 2020

BT-TOP-23 A1

BLUTEST LABORATORIES LIMITED, 5 Robroyston Oval, Nova Business Park, Glasgow, UK, G33 1AP
Telephone: +44 (0)141 558 2732. Email: info@blutest.com. Web site: www.blutest.com.
Company Registration Number: SC364409 VAT Registration Number: GB 979 1131 95 UKAS Number: 4597

Efficacy Test Report Certificate



PROTOCOL SUMMARY

The basic virucidal efficacy test is set up with three concentrations of test product solution and a 2-minute contact time. Virus is exposed to disinfectant in 24-well plates, then neutralised, serially diluted and virus titred in 96-well tissue culture plates to determine the tissue culture infectious dose₅₀ (TCID₅₀) of surviving virus. *Vaccinia virus* VR-1549 Elstree strain / Vero cells are assayed in parallel in each test. TCID₅₀ is determined by the method of Karber¹.

Cytotoxicity control

The test product solution is measured for its effects on the host cells used to propagate the virus, to determine the sensitivity of the assay.

Interference control

The effect of the cells after treatment of the test product solution are verified to ensure the cells can show susceptibility for virus infection. This is compared against cells that have not been treated with test product.

Disinfectant suppression control VS1

Virus is added to the highest concentration of test product solution and then the mixture immediately removed and neutralised. The neutralised virus titre is then determined to assess the efficiency of the neutralisation procedure.

Disinfectant suppression control VS2

Internal control which adds virus to neutralised test product solution to assess the efficiency of the neutralisation procedure.

No column Control

Internal control on the highest contact time to assess any impact of the Microspin™ S 400 HR columns.

Virus recovery control

Virus titre is determined for virus in contact with sterile distilled water at t=0, t = 2 and at t =15. The virus titre after 2 minutes is then compared to the recovery of disinfectant-treated virus to measure the log reduction in virus titre. The virus titre at 15 minutes is compared to the reference virus inactivation control.

Reference virus inactivation control

Virus is exposed to 0.7% W/V formaldehyde and the recovery of virus determined by TCID₅₀ after 5 and 15 minutes, in order to assess that the test virus has retained reproducible biocide resistance. In addition, the formaldehyde cytotoxicity of neutralised formaldehyde is determined, to measure assay sensitivity.

1Kärber, G.: Beitrag zur Kollektiven Behandlung Pharmakologischer Reihenversuche. Arch. Exp. Path. Pharmacol. 162 (1931): 480-487.

Page 2 of 7

SOP 11000
SOP 8003 EN14476 Vaccinia REPORT TEMPLATE V01
Effective Date: 23 March 2020

BT-TOP-23 A1

BLUTEST LABORATORIES LIMITED, 5 Robroyston Oval, Nova Business Park, Glasgow, UK, G3 7 1AP
Telephone: +44 (0)141 958 2782. Email: info@blutest.com. Web site: www.blutest.com.
Company Registration Number: SC364409 VAT Registration Number: GB 979 1131 96 UKAS Number: 4597



Vaccinia virus (VR-1549) Elstree strain Test Results

Test Results					
Concentration	10.0% (w/v)		50.0% (w/v)		80.0% (w/v)
Exposure Time t = 2 mins	data	TCID ₅₀ /ml	data	TCID ₅₀ /ml	TCID ₅₀ /ml
Raw Data	3.00	3.16E+04	1.00	3.16E+02	3.16E+01
log	666000	3.16E+04	600000	3.16E+02	3.16E+01
log difference		4.50		2.50	1.50
		2.00		4.00	5.00

EN14476:2013 + A2:2019 Suspension test for the efficacy of Alcohol Free BT-57, BT-TOP-23 from Top Dental Products Limited against Vaccinia virus VR-1549 under DIRTY conditions

Summary Table									
Product:	Interfering substance	Concentration	Level of cytotoxicity	lg TCID ₅₀					>4 lg reduction after 'X' Min
				0 min	5 min	15 min	30 min	60 min	
Alcohol Free BT-57	3.0g/l BSA + 3.0ml/l erythrocytes	80.0% (v/v)	1.50	2.33	1.50	n.a.	n.a.	n.a.	<2 mins
		50.0% (v/v)	1.50	n.a.	2.50	n.a.	n.a.	n.a.	2 mins
		10.0% (v/v)	1.50	n.a.	4.50	n.a.	n.a.	n.a.	>2 mins
Alcohol Free BT-57	3.0g/l BSA	80.0% (v/v)	1.50	n.a.	1.50	n.a.	n.a.	n.a.	<2 mins
		50.0% (v/v)	1.50	n.a.	2.50	n.a.	n.a.	n.a.	2 mins
		10.0% (v/v)	1.50	n.a.	4.50	n.a.	n.a.	n.a.	>2 mins
Virus Control	DIRTY			6.33	6.50	6.17	n.a.	n.a.	n.a.
Virus Control	CLEAN			6.33	6.50	6.33	n.a.	n.a.	n.a.
Formaldehyde	PBS	0.7% (w/v)	2.50				5 min	15 min	>15 mins
							4.67	2.50	>15 mins

SOP 1.000
SOP 8003 EN14476 Vaccinia REPORT TEMPLATE V01
Effective Date: 23 March 2020

BT-TOP-23 A1

Efficacy Test Report Certificate



Vaccinia virus (VR-1549) Elstree strain Control Data

EN14476:2013 + A2:2019 Suspension test for the efficacy of Alcohol Free BT-57, BT-TOP-23 from Top Dental Products Limited against Vaccinia virus VR-1549 under DIRTY conditions

Controls									
Virus Recovery 0 min		Virus Recovery 2 min		Virus Recovery 15 min		Cytotoxicity		Disinfectant Suppression VS	
raw data	TCID ₅₀ /ml	raw data	TCID ₅₀ /ml	raw data	TCID ₅₀ /ml	raw data	TCID ₅₀ /ml	raw data	TCID ₅₀ /ml
4.83	2.15E+06	5.00	3.16E+06	4.67	1.47E+06	0.00	3.16E+01	0.83	2.15E+02
666650	2.15E+06	666650	3.16E+06	666640	1.47E+06	000000	3.16E+01	311000	2.15E+02
	6.33		6.50		6.17		1.50	2.33	6.67
								4.17	-0.17

Formaldehyde reference inactivation controls									
Cytotoxicity		Exposure time		0.7% Formaldehyde		5 mins		15 mins	
raw data	TCID ₅₀ /ml	raw data	TCID ₅₀ /ml	raw data	TCID ₅₀ /ml	raw data	TCID ₅₀ /ml	raw data	TCID ₅₀ /ml
1.00	3.16E+02		3.17	4.64E+04	1.00	3.16E+02	3.16E+02		
600000	3.16E+02		666100	4.64E+04	600000	3.16E+02	3.16E+02		
	2.50	log difference		4.67	2.50		3.67		

No column Control									
raw data		TCID ₅₀ /ml		raw data		TCID ₅₀ /ml		raw data	
5.17		4.64E+06		5.17		4.64E+06		5.17	
666661		4.64E+06		666661		4.64E+06		666661	
		6.67				6.67			

Stock Virus (TCID ₅₀)									
raw data		TCID ₅₀ /ml		raw data		TCID ₅₀ /ml		raw data	
6.33		6.76E+07		6.33		6.76E+07		6.33	
666662000		6.76E+07		666662000		6.76E+07		666662000	

Efficacy Test Report Certificate



Vaccinia virus (VR-1549) Elstree strain Control Data

Parallel Control Test											
Controls				Test Results							
Virus Recovery 0 min		Virus Recovery 2 min		Virus Recovery 15 min		10.0% (v/v)		50.0% (v/v)		80.0% (v/v)	
raw data	TCD ₅₀ /ml	raw data	TCD ₅₀ /ml	raw data	TCD ₅₀ /ml	Exposure Time		data	TCD ₅₀ /ml	data	TCD ₅₀ /ml
4.83	2.15E+06	5.00	3.16E+06	4.83	2.15E+06	t = 2 mins		3.00	3.16E+04	1.00	3.16E+01
666650	2.15E+06	666660	3.16E+06	666650	2.15E+06	Raw data		664200	3.16E+04	600000	3.16E+01
	6.33		6.50		6.33	log			4.50		2.50
						log difference			2.00		4.00
									2.00		5.00



CONCLUSION

Verification of the methodology

A test is only valid if the following criteria are fulfilled:

- The titre of the test suspension of at least 10^8 TCID₅₀/ml is sufficiently high to at least enable a titre reduction of 4 lg to verify the method.
- Detectable titre reduction is at least 4 log₁₀.
- Difference of the logarithmic titre of the virus control minus the logarithmic titre of the test virus in the reference inactivation test is between:
 - Between 0.75 and 3.5 after 5 min and between 2.0 and 4.0 after 15 min for Vaccinia virus
- Cytotoxicity of the product solution does not affect cell morphology and growth or susceptibility for the test virus in the dilutions of the test mixtures which are necessary to demonstrate a 4 log₁₀ reduction of the virus.
- The interference control result does not show a difference of > 1.0 log₁₀ of virus titre for test product treated cells in comparison to the non-treated cells.
- Neutralisation validation. This is called the disinfectant suppression test in this protocol. The disinfectant was neutralised by column chromatography through an Illustra Microspin S-400 HR column to achieve the best possible neutralisation available for this test. The difference for virus is greater than 0.5 log₁₀ indicating rapid irreversible virucidal activity of the disinfectant by dilution at a concentration of 80.0% v/v for VS1. This neutralisation validation has been verified by VS2, which shows the product has been successfully neutralised.

According to EN 14476:2013 + A2:2019, **Alcohol Free BT-57 POSSESSES VIRUCIDAL** activity at a concentration of **80.0% v/v** of the working concentration as tested after **2 MINUTES** at **20°C** under **DIRTY** conditions (3.0 g/l bovine albumin + 3.0 ml/l erythrocytes) against *Vaccinia virus* VR-1549 Elstree strain / Vero cells.

This product therefore is effective against all enveloped viruses as defined in EN 14476:2013 + A2:2019 Annex A*. This therefore includes all coronaviruses and SARS-CoV-2.

Authorized signatory

Mr.

Dr Chris Woodall, Director
BluTest Laboratories Ltd
Glasgow, UK.
Date: 27 AUGUST 2020

DISCLAIMER

The results in this test report only pertain to the sample supplied.

BLU Testing (BT) has performed the testing detailed in this report using reasonable skill and care and has used reasonable endeavours to carry out the testing in accordance with an EN 14476 protocol. All forecasts, recommendations and results contained in this report are submitted in good faith. However, other than as expressly set out in this report, no warranty is given (i) in relation to the testing or the use(s) to which any results or deliverables produced in the course of the testing are or may be put by the Client or their fitness or suitability for any particular purpose or under any special conditions notwithstanding that any such purpose or conditions may have been made known to BT or (ii) that the intended results or deliverables from the testing can be achieved or (iii) that the Client can freely make use of the results or the deliverables without infringing any third party intellectual property rights and the Client will be deemed to have satisfied itself in this regard. BT shall have no liability (which is hereby excluded to the fullest extent permissible by law) in respect of any loss, liability or damage, including without limitation any indirect and/or consequential loss such as loss of profit or loss of business, market or goodwill, that the Client may suffer directly or indirectly as a result of or in connection with: (i) the performance of the testing; (ii) the use of any materials, samples or other information provided by the Client for use in the testing; and (iii) the Client's reliance upon or use of any results or deliverables provided as part of the testing.

Amendment 1: BT-TOP-23 EN14476 Vaccinia Report 25 Aug 20 | M CW: Batch number added | M 27 August 2020

Page 6 of 7

SOP 11000
SOP 8003 EN14476 Vaccinia REPORT TEMPLATE V01
Effective Date: 23 March 2020

HT-TOP-23 A1

BLUTEST LABORATORIES LIMITED, 5 Robroydon Oval, Nova Business Park, Glasgow, UK, G68 1AP
Telephone: +44 (0)141 558 2782. Email: info@blutest.com. Web site: www.blutest.com.
Company Registration Number: SC364409 VAT Registration Number: GB 979 3131 95 UKAS Number: 4592

Efficacy Test Report Certificate



*EN 14476 2013 + A2 2019 Annex A (informative – Enveloped viruses)

Poxviridae
Herpesviridae
Filoviridae (e.g. Ebola, Marburg)
Flavivirus
Hepatitis C Virus (HCV)
Hepatitis Delta Virus (HDV)
Influenza Virus
Paramyxoviridae
Rubella Virus
Measles Virus
Rabies Virus
Coronavirus (e.g. SARS, MERS)
Human Immunodeficiency Virus (HIV)
Human T Cell Leukemia Virus (HTLV)
Hepatitis B virus (HBV)

Reference: Van Regenmortel MHV et al., Eds.: Virus Taxonomy, Classification and Nomenclature of Viruses, seventh report of the international committee on taxonomy of viruses. Academic Press, San Diego, 2000

Page 7 of 7

SOP 11000
SOP 8003 EN14476 Vaccinia REPORT TEMPLATE V01
Effective Date: 23 March 2020

BT-TOP-33 A1

BLUTEST LABORATORIES LIMITED, 5 Robroyston Oval, Nova Business Park, Glasgow, UK, G33 1AP
Telephone: +44 (0)141 558 2782. Email: info@blutest.com. Web site: www.blutest.com
Company Registration Number: SC364409 VAT Registration Number: GB 979 1131 96 UKAS Number: 4597

